

**THE
UNIVERSITY
OF RHODE ISLAND**

**DIVISION OF
ADMINISTRATION
AND FINANCE**

THINK BIG  WE DO™

PURCHASING DEPARTMENT

10 Tootell Road, Suite 3, Kingston, RI 02881 USA

p: 401.874.2171

f: 401.874.2306

uri.edu/purchasing



DATE: 5/19/16

ADDENDUM #2

BID NO.:100249

OPENING: 5/26/16 @ 3:00pm

COMMODITY: Dual User Glove Box

This Addendum 2 is the result of a question received by the final due date for all Questions regarding this bid#100249. The question and the corresponding answer follows:

Question #1: On page 2, item 1, position h; Moisture sensor must be electrochemical. Is this correct? Normally The O2 sensor is electrochemical or solid-state designed and the H2O (moisture) is electrolytic in design.

Answer #1: The line reading "moisture control must be electrochemical" should read "oxygen sensor should be electrochemical".

If you have already submitted a bid and need to make changes, based on the information within the Addendum, please submit a new bid response and indicate that this submission supersedes the prior.



Debra Lee
URI Assistant Purchasing Agent
Rev. 9-1-15